



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

AM-S-Y2304

Biological matrix : Serum

Date of publication :
Expiration date : 2028-04-19

MATERIAL NUMBER	ANALYTE GROUPING	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
AM-S-Y2304	Perfluorinated compounds	PFOA	2.32	µg/L	0.0460	16
		PFOS	10.5	µg/L	0.488	18
		PFHxS	1.98	µg/L	0.0569	20
		PFHxA	0.873	µg/L	0.0214	18
		PFNA	0.640	µg/L	0.0194	20
		PFUdA	1.64	µg/L	0.0388	19
		PFDA	3.33	µg/L	0.0942	19
		PFHpS	1.31	µg/L	0.0509	14
		PFHpA	0.446	µg/L	0.0196	16
	PFBA PFBS	PFBA	0.412	µg/L	0.0184	9
		PFBS	0.577	µg/L	0.0191	17
	PFOA BL	PFOA branch	0.111	µg/L	0.0453	4
		PFOA linear	2.31	µg/L	0.0568	10
	PFOS branch	PFOS branch	3.44	µg/L	0.349	11
	PFOS linear	PFOS linear	6.82	µg/L	0.323	14

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales