



Centre de toxicologie / INSPQ

# Biological Reference Material - Evaluation Report -

**PC-U-N2310**

**Biological matrix :** Urine  
**Date of publication :**  
**Expiration date :** 2028-05-17

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
PC-U-N2310	Urine	Selenium	131	µg/L	11.0	6

\* Standard uncertainty ( $\mu$ x): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada  
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : [ctq@inspq.qc.ca](mailto:ctq@inspq.qc.ca) | Site Web : [www.inspq.qc.ca/ctq](http://www.inspq.qc.ca/ctq)  
Magasin virtuel : [www.inspq.qc.ca/ctq/vente](http://www.inspq.qc.ca/ctq/vente) | Virtual store/reference material : [www.inspq.qc.ca/ctq/sales](http://www.inspq.qc.ca/ctq/sales)