



Centre de toxicologie / INSPQ

# Biological Reference Material

## - Evaluation Report -

**PC-B-M2313**

**Biological matrix :** Blood

**Date of publication :**

**Expiration date :** 2027-09-20

| MATERIAL<br>NUMBER | Biological matrix | ANALYTE | ASSIGNED<br>VALUE | UNIT | STANDARD<br>UNCERTAINTY * | N  |
|--------------------|-------------------|---------|-------------------|------|---------------------------|----|
| PC-B-M2313         | Blood             | Mercury | 4.74              | µg/L | 0.0527                    | 15 |

\* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada  
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : [ctq@inspq.qc.ca](mailto:ctq@inspq.qc.ca) | Site Web : [www.inspq.qc.ca/ctq](http://www.inspq.qc.ca/ctq)  
Magasin virtuel : [www.inspq.qc.ca/ctq/vente](http://www.inspq.qc.ca/ctq/vente) | Virtual store/reference material : [www.inspq.qc.ca/ctq/sales](http://www.inspq.qc.ca/ctq/sales)