



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

PC-U-H2405

Biological matrix : Urine

Date of publication :

Expiration date : 2029-01-08

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
PC-U-H2405	Urine	Mercury	55.4	µg/L	1.61	17

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales