



Centre de toxicologie / INSPQ

# Biological Reference Material

## - Evaluation Report -

**OS-U-P2502**

Biological matrix : Urine

Date of publication :  
Expiration date : 2030-03-31

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-P2502	Urine	2,4-Dichlorophenol	1.87	µg/L	0.162	3
		2,4-Dichlorophenoxyacetic acid (2,4-D)	1.29	µg/L	0.0616	5
		2,5-Dichlorophenol	1.28	µg/L	0.0770	3
		5-OH-imidacloprid	0.680	µg/L	0.383	5
		Acetamiprid	0.353	µg/L	0.0433	5
		AMPA	2.47	µg/L	0.0767	8
		Clothianidin	1.68	µg/L	0.0530	5
		Desmethyl-acetamiprid	0.888	µg/L	0.0857	4
		Desmethyl-clothianidin	0.492	µg/L	0.424	3
		Dinotefuran	2.57	µg/L	0.705	4
		Glyphosate	2.08	µg/L	0.0571	8
		Sulfoxaflor	0.309	µg/L	0.0236	4
		Thiacloprid	0.612	µg/L	0.00856	4
		Thiamethoxam	0.923	µg/L	0.278	3
Triclosan	232	µg/L	11.2	7		

\* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada  
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq  
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales