



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

OS-U-B2508

Biological matrix : Urine

Date of publication :

Expiration date :

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-B2508	Urine	Bisphenol A	1.56	µg/L	0.0728	8
		Bisphenol B	0.627	µg/L	0.152	4
		Bisphenol F	2.97	µg/L	0.383	8
		Bisphenol S	2.05	µg/L	0.181	9
		Bisphenol Z	0.449	µg/L	0.0780	4
	Urine	Bis(1,3-dichloro-2-propyl) phosphate (BDGP)	2.60	µg/L	0.181	6
		Bis(1-chloro-2-propyl) phosphate (BCPP)	1.40	µg/L	0.208	5
		Bis(2-chloroethyle) phosphate (BCEtP)	0.602	µg/L	0.00445	4
		Diphenyl phosphate (DPhP)	3.41	µg/L	0.132	6

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales