



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

OS-U-P2505

Biological matrix : Urine

Date of publication :

Expiration date :

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-P2505	Urine	2,4-Dichlorophenoxyacetic acid (2,4-D)	2.16	µg/L	0.0865	6
		5-OH-imidacloprid	0.535	µg/L	0.309	4
		Acetamiprid	0.960	µg/L	0.0747	6
		AMPA	1.24	µg/L	0.0845	8
		Clothianidin	3.36	µg/L	0.675	6
		Desmethyl-acetamiprid	0.467	µg/L	0.0678	5
		Desmethyl-clothianidin	5.03	µg/L	2.89	4
		Desmethyl-thiamethoxam	1.29	µg/L	0.155	3
		Dinotefuran	1.00	µg/L	0.175	5
		Glyphosate	2.25	µg/L	0.100	8
		Sulfoxaflor	0.492	µg/L	0.0373	4
		Thiacloprid	1.23	µg/L	0.00835	5
		Thiamethoxam	0.377	µg/L	0.157	4
		Triclosan	471	µg/L	45.5	6

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales