



Centre de toxicologie / INSPQ

# Biological Reference Material

## - Evaluation Report -

**PC-S-A2603**

**Biological matrix :** Serum

**Date of publication :**

**Expiration date :**

2030-10-14

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
PC-S-A2603	Serum	Aluminum	17.2	µg/L	1.13	7

\* Standard uncertainty ( $\mu$ x): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada  
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : [ctq@inspq.qc.ca](mailto:ctq@inspq.qc.ca) | Site Web : [www.inspq.qc.ca/ctq](http://www.inspq.qc.ca/ctq)  
Magasin virtuel : [www.inspq.qc.ca/ctq/vente](http://www.inspq.qc.ca/ctq/vente) | Virtual store/reference material : [www.inspq.qc.ca/ctq/sales](http://www.inspq.qc.ca/ctq/sales)