



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

PC-U-R2603

Biological matrix : Urine
Date of publication :
Expiration date : 2030-09-19

| MATERIAL NUMBER | Biological matrix | ANALYTE | ASSIGNED VALUE | UNIT | STANDARD UNCERTAINTY * | N |
|-----------------|-------------------|---------|----------------|------|------------------------|----|
| PC-U-R2603 | Urine | Copper | 458 | µg/L | 9.88 | 13 |
| | Urine | Zinc | 545 | µg/L | 11.2 | 13 |

* Standard uncertainty (µ): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales