



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

PC-S-E2606

Biological matrix : Serum

Date of publication :
Expiration date : 2031-01-12

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
PC-S-E2606	Serum	Copper	1700	µg/L	33.3	19
	Serum	Manganese	0.623	µg/L	0.0376	6
	Serum	Selenium	232	µg/L	4.25	18
	Serum	Zinc	1750	µg/L	23.4	20

* Standard uncertainty (μ x): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales