



Centre de toxicologie / INSPQ

# Biological Reference Material

## - Evaluation Report -

**PC-U-R2606**

**Biological matrix :** Urine  
**Date of publication :**  
**Expiration date :** 2031-01-07

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
PC-U-R2606	Urine	Copper	63.8	µg/L	0.587	15
	Urine	Zinc	463	µg/L	4.30	14

\* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada  
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq  
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales