



Centre de toxicologie / INSPQ

# Biological Reference Material

## - Evaluation Report -

**PC-U-S2606**

Biological matrix : Urine

Date of publication :

Expiration date :

2030-11-10

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
PC-U-S2606	Urine	Arsenic Arsenobetaine	0.731	µg/L	0.0385	4
	Urine	Arsenic DMAA	1.58	µg/L	0.0680	6
	Urine	Arsenic MAA	68.4	µg/L	1.71	7
	Urine	Inorganic arsenic	70.7	µg/L	1.45	14
	Urine	Total arsenic	70.1	µg/L	1.26	17

\* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada  
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : [ctq@inspq.qc.ca](mailto:ctq@inspq.qc.ca) | Site Web : [www.inspq.qc.ca/ctq](http://www.inspq.qc.ca/ctq)  
Magasin virtuel : [www.inspq.qc.ca/ctq/vente](http://www.inspq.qc.ca/ctq/vente) | Virtual store/reference material : [www.inspq.qc.ca/ctq/sales](http://www.inspq.qc.ca/ctq/sales)