



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

OS-U-B2602

Biological matrix : Urine

Date of publication :

Expiration date :

2031-04-17

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-B2602	Urine	Bisphenol A	5.43	µg/L	0.766	8
		Bisphenol B	1.77	µg/L	0.188	4
		Bisphenol F	4.68	µg/L	0.274	7
		Bisphenol S	0.617	µg/L	0.0165	7
		Bisphenol Z	4.96	µg/L	0.419	4
	Urine	Bis(1,3-dichloro-2-propyl) phosphate (BDGP)	3.94	µg/L	0.176	8
		Bis(1-chloro-2-propyl) phosphate (BCPP)	1.65	µg/L	0.129	7
		Bis(2-chloroethyle) phosphate (BCETP)	0.961	µg/L	0.105	6
		Diphenyl phosphate (DPHP)	30.3	µg/L	2.81	8

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales