



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

OS-U-B2604

Biological matrix : Urine

Date of publication :

Expiration date :

2031-04-23

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-B2604	Urine	Bisphenol A	0.962	µg/L	0.0457	8
		Bisphenol B	1.47	µg/L	0.136	4
		Bisphenol F	5.05	µg/L	0.340	7
		Bisphenol S	6.12	µg/L	0.339	7
		Bisphenol Z	1.61	µg/L	0.102	4
	Urine	Bis(1,3-dichloro-2-propyl) phosphate (BDGP)	6.84	µg/L	0.270	8
		Bis(1-chloro-2-propyl) phosphate (BCPP)	2.92	µg/L	0.445	7
		Bis(2-chloroethyle) phosphate (BCEtP)	1.27	µg/L	0.209	6
		Diphenyl phosphate (DPhP)	18.0	µg/L	0.917	8

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada
Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq
Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales