



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

OS-U-P2601

Biological matrix : Urine

Date of publication :

Expiration date :

2031-04-17

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-P2601	Urine	2,4-Dichlorophenoxyacetic acid (2,4-D)	0.361	µg/L	0.0127	4
		5-OH-imidacloprid	0.267	µg/L	0.154	3
		AMPA	0.314	µg/L	0.0524	6
		Clothianidin	0.657	µg/L	0.0175	4
		Desmethyl-acetamiprid	0.623	µg/L	0.0135	3
		Desmethyl-clothianidin	0.657	µg/L	0.565	3
		Glyphosate	0.157	µg/L	0.0194	6
		Thiamethoxam	0.683	µg/L	0.0405	3
		Triclosan	3.34	µg/L	0.186	5

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada

Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq

Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales