



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

OS-U-P2602

Biological matrix : Urine

Date of publication :

Expiration date :

2031-05-01

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-P2602	Urine	2,4-Dichlorophenoxyacetic acid (2,4-D)	1.06	µg/L	0.125	5
		5-OH-imidacloprid	0.683	µg/L	0.383	3
		Acetamiprid	1.54	µg/L	0.121	4
		AMPA	27.6	µg/L	1.47	7
		Clothianidin	3.80	µg/L	0.236	4
		Desmethyl-acetamiprid	1.84	µg/L	0.198	3
		Desmethyl-clothianidin	0.431	µg/L	0.354	3
		Glyphosate	8.01	µg/L	0.366	8
		Sulfoxaflor	4.00	µg/L	0.0946	3
		Thiacloprid	0.242	µg/L	0.0147	3
		Thiamethoxam	1.33	µg/L	0.441	3
		Triclosan	81.4	µg/L	3.53	5

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada

Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq

Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales