



Centre de toxicologie / INSPQ

Biological Reference Material

- Evaluation Report -

OS-U-P2603

Biological matrix : Urine

Date of publication :

Expiration date :

2031-05-01

MATERIAL NUMBER	Biological matrix	ANALYTE	ASSIGNED VALUE	UNIT	STANDARD UNCERTAINTY *	N
OS-U-P2603	Urine	2,4-Dichlorophenoxyacetic acid (2,4-D)	2.41	µg/L	0.173	5
		Acetamiprid	2.11	µg/L	0.114	4
		AMPA	4.03	µg/L	0.241	7
		Clothianidin	0.797	µg/L	0.0307	4
		Desmethyl-acetamiprid	4.10	µg/L	0.447	3
		Desmethyl-clothianidin	1.05	µg/L	0.843	3
		Glyphosate	19.1	µg/L	1.35	8
		Sulfoxaflor	3.20	µg/L	0.0405	3
		Thiacloprid	2.09	µg/L	0.281	3
		Triclosan	31.6	µg/L	0.111	5

* Standard uncertainty (µx): characteristic of the dispersion of results used to calculate the assigned value; expressed in the same format as the standard deviation.

Centre de toxicologie / INSPQ | 945, ave. Wolfe | Québec (Québec) G1V 5B3 | Canada

Tél. : (418) 650-5115, poste 4254 | Fax : (418) 654-2148 | Courriel : ctq@inspq.qc.ca | Site Web : www.inspq.qc.ca/ctq

Magasin virtuel : www.inspq.qc.ca/ctq/vente | Virtual store/reference material : www.inspq.qc.ca/ctq/sales